

Hengst SE
Nienkamp 55-85 · 48147 Münster
Postfach 20 14 54 · 48095 Münster
Germany

T +49 (0)251 20202-0
F +49 (0)251 20202-900

info@hengst.de
www.hengst.de

Product testing MNS 07 according to DIN EN 14683

According to the investigations by SGS-CSTC Standards Technical Services CO., LTD. we confirm that you have passed the following tests:


- Bacterial filtration efficiency: $\geq 98 \%$
- Differential pressure: $< 60 \text{ Pa/cm}^2$
- Splash resistance: $\geq 16 \text{ kPa}$

The tested probe does fulfil the requirements of DIN EN 14683 for medical masks type II R.

Münster, 07/14/2020



Christopher Heine
CEO



ppa. Oliver Nabrotzky
Group Vice President